

Report No.:

Test Time: 2025-10-08 15:08

Luminaire Property

Luminaire Manufacturer:

Luminaire Description: 17D

Luminous Width (mm): -96

Current: 0.174 A

Power Factor: 0.982

Luminous Length (mm): -96

Voltage: 225.8 V

Power: 38.60 W

Photometric Results

CIE Class: Direct

Measurement Flux: 4936.1 lm

Downward Ratio: 100%

Horizontal Diffuse Angle(50%): H22.2

Vertical Diffuse Angle(50%): V22.2

Luminaire Efficacy Rating (LER): 127.93

Max. Intensity: 19547.91 cd

S/MH(C0/C180): 0.38

Total Rated Lamp Lumens: 4936.1 lm

Efficiency: 100%


Upward Ratio: 0%

Central Intensity: 19547.9 cd


Pos of Max. Intensity: H0 V0

S/MH(C90/C270): 0.38

Picture Of Luminaire



Luminous Intensity Distribution Curve



— C0-C180 — C90-C270

C Plane (°):0.0-180.0: 180.0

Test Lab:

Test Type: TYPE C

Temperature:

Operator:

Gamma Plane (°):0.0-90.0:1.0



Test Device: CHL-6E

Distance: 2.905 m

Humidity:

Inspector:


Luminous Intensity Distribution Curve



C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:


Luminous Intensity Distribution Curve(cd/klm)



C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Isocandela (rectangle)



Imax (100%): 19548 cd

(10%): 1955 cd	(16%): 3128 cd
(25%): 4887 cd	(40%): 7819 cd
(63%): 12315 cd	(100%): 19548 cd

C Plane (°):0.0-180.0: 180.0

Test Lab:

Test Type: TYPE C

Temperature:

Operator:

Gamma Plane (°):0.0-90.0:1.0


Test Device: CHL-6E

Distance: 2.905 m

Humidity:

Inspector:

IsoLux Plot




Mounting Height: 10.0m Max Lux(100%): 195.5 lx

(1%): 2.0 lx	(2%): 3.9 lx
(5%): 9.8 lx	(10%): 19.5 lx
(20%): 39.1 lx	(50%): 97.7 lx
(100%): 195.5 lx	


C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Illuminance at a Distance



Vertical IsoLux Plot




Lowest(m): 1.0m Highest(m): 11.0m Max Lux: 19547.9 lx
— (1%): 195.5 lx — (2%): 391.0 lx
— (5%): 977.4 lx — (10%):1954.8 lx
— (20%):3909.6 lx — (50%):9774.0 lx
— (100%):19547.9 lx

C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

The Average Illuminance Effective Figure



UGR Table


Reflectance:										
Ceiling (cavity)	0.7	0.7	0.5	0.5	0.3	0.7	0.7	0.5	0.5	0.3
Wall	0.5	0.3	0.5	0.3	0.3	0.5	0.3	0.5	0.3	0.3
Reference plane	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2
Room dimensions	Viewed crosswise					Viewed endwise				
X=2H Y=2H	19.9	20.6	20.2	20.8	21.0	19.8	20.5	20.0	20.7	20.9
3H	19.8	20.4	20.1	20.6	20.9	19.6	20.3	19.9	20.5	20.7
4H	19.7	20.3	20.0	20.5	20.8	19.6	20.2	19.9	20.4	20.7
6H	19.6	20.2	20.0	20.5	20.7	19.5	20.0	19.8	20.3	20.6
8H	19.6	20.1	19.9	20.4	20.7	19.5	20.0	19.8	20.3	20.6
12H	19.6	20.1	19.9	20.4	20.7	19.4	19.9	19.8	20.2	20.5
X=4H Y=2H	19.7	20.3	20.0	20.5	20.8	19.6	20.2	19.9	20.4	20.7
3H	19.6	20.1	19.9	20.4	20.7	19.4	19.9	19.8	20.2	20.5
4H	19.5	19.9	19.8	20.3	20.6	19.3	19.8	19.7	20.1	20.5
6H	19.4	19.8	19.8	20.1	20.5	19.3	19.6	19.7	20.0	20.4
8H	19.3	19.7	19.8	20.1	20.5	19.2	19.6	19.6	20.0	20.4
12H	19.3	19.6	19.7	20.0	20.5	19.2	19.5	19.6	19.9	20.3
X=8H Y=4H	19.3	19.7	19.8	20.1	20.5	19.2	19.6	19.6	20.0	20.4
6H	19.2	19.5	19.7	20.0	20.4	19.1	19.4	19.6	19.8	20.3
8H	19.2	19.4	19.7	19.9	20.4	19.1	19.3	19.5	19.8	20.2
12H	19.1	19.4	19.6	19.8	20.3	19.0	19.2	19.5	19.7	20.2
X=12H Y=4H	19.3	19.6	19.7	20.0	20.5	19.2	19.5	19.6	19.9	20.3
6H	19.2	19.4	19.7	19.9	20.4	19.1	19.3	19.5	19.8	20.2
8H	19.1	19.4	19.6	19.8	20.3	19.0	19.2	19.5	19.7	20.2
Variations with the observer position at spacings:										
S=1.0H	+5.6/-1.\$					+5.6/-1.\$				
S=1.5H	+8.4/-1.\$					+8.4/-1.\$				
S=2.0H	+10.4/-1.\$					+10.4/-1.\$				

Calculate in accordance with CIE Pub.117. The table is revised with 4936lm ($8\log(F/F_0) = 5.5$).

C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

FLUX DISTRIBUTION TABLE BASED ON THE IESNA LUMINAIRE CLASSIFICATION SYSTEM



ZONE	LUMENS	% LAMP LUMENS
FORWARD LIGHT	2488	50.4
FL (0°-30°)	2096	42.5
FM (30°-60°)	391	7.9
FH (60°-80°)	0	0.0
FVH (80°-90°)	0	0.0
BACK LIGHT	2449	49.6
BL (0°-30°)	2074	42.0
BM (30°-60°)	375	7.6
BH (60°-80°)	0	0.0
BVH (80°-90°)	0	0.0
UP LIGHT	0	0.0
UL (90°-100°)	0	0.0
UH (100°-180°)	0	0.0
TRAPPED LIGHT	NA	NA

BUG(Backlight,Uplight,Glare) Rating Base On TM-15-07	
Asymmetrical Luminaire Types (Type I,II,III,IV)	B3 U0 G0
Quadrilateral Symmetrical Luminaire Types (Type V,Area Light)	B3 U0 G0

C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Zonal Lumen

Gamma [°]	I _{mean} [cd]	Zonal Flux [lm]	Sum Zonal Flux [lm]	Rel Zonal Flux [%]	Sum Rel Zonal Flux [%]
0.0-1.0	19528.2	18.7	18.7	0.38	0.38
1.0-2.0	19435.9	55.8	74.5	1.13	1.51
2.0-3.0	19199.3	91.8	166.3	1.86	3.37
3.0-4.0	18723.6	125.3	291.7	2.54	5.91
4.0-5.0	17959.6	154.5	446.2	3.13	9.04
5.0-6.0	16978.0	178.4	624.6	3.62	12.65
6.0-7.0	15857.7	196.9	821.5	3.99	16.64
7.0-8.0	14625.8	209.3	1030.8	4.24	20.88
8.0-9.0	13303.7	215.6	1246.5	4.37	25.25
9.0-10.0	11929.6	215.9	1462.4	4.37	29.63
10.0-11.0	10569.1	211.2	1673.6	4.28	33.91
11.0-12.0	9276.8	202.8	1876.4	4.11	38.01
12.0-13.0	8084.6	191.9	2068.3	3.89	41.90
13.0-14.0	7024.3	179.8	2248.1	3.64	45.55
14.0-15.0	6116.4	167.9	2416.1	3.40	48.95
15.0-16.0	5355.0	156.9	2573.0	3.18	52.13
16.0-17.0	4717.1	146.9	2719.9	2.98	55.10
17.0-18.0	4182.7	137.9	2857.8	2.79	57.90
18.0-19.0	3733.3	129.9	2987.8	2.63	60.53
19.0-20.0	3363.0	123.1	3110.9	2.49	63.02
20.0-21.0	3065.9	117.7	3228.6	2.39	65.41
21.0-22.0	2822.6	113.4	3342.0	2.30	67.71
22.0-23.0	2619.4	109.9	3452.0	2.23	69.93
23.0-24.0	2450.2	107.1	3559.1	2.17	72.10
24.0-25.0	2308.7	105.0	3664.1	2.13	74.23
25.0-26.0	2186.8	103.2	3767.3	2.09	76.32
26.0-27.0	2083.3	101.9	3869.3	2.07	78.39
27.0-28.0	1994.8	101.0	3970.3	2.05	80.43
28.0-29.0	1916.2	100.3	4070.5	2.03	82.47
29.0-30.0	1840.3	99.4	4169.9	2.01	84.48
30.0-31.0	1750.8	97.4	4267.4	1.97	86.45
31.0-32.0	1638.1	93.9	4361.2	1.90	88.35
32.0-33.0	1507.5	88.8	4450.1	1.80	90.15
33.0-34.0	1367.1	82.7	4532.8	1.68	91.83
34.0-35.0	1219.6	75.8	4608.6	1.53	93.36
35.0-36.0	1070.1	68.1	4676.7	1.38	94.75

C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Zonal Lumen (Continue 1)

Gamma [°]	I _{mean} [cd]	Zonal Flux [lm]	Sum Zonal Flux [lm]	Rel Zonal Flux [%]	Sum Rel Zonal Flux [%]
36.0-37.0	923.7	60.3	4736.9	1.22	95.97
37.0-38.0	778.3	52.0	4788.9	1.05	97.02
38.0-39.0	633.7	43.3	4832.2	0.88	97.89
39.0-40.0	494.7	34.5	4866.7	0.70	98.59
40.0-41.0	365.5	26.0	4892.7	0.53	99.12
41.0-42.0	250.7	18.2	4910.9	0.37	99.49
42.0-43.0	154.5	11.4	4922.4	0.23	99.72
43.0-44.0	83.9	6.3	4928.7	0.13	99.85
44.0-45.0	45.3	3.5	4932.2	0.07	99.92
45.0-46.0	26.0	2.0	4934.2	0.04	99.96
46.0-47.0	12.3	1.0	4935.2	0.02	99.98
47.0-48.0	4.6	0.4	4935.6	0.01	99.99
48.0-49.0	2.8	0.2	4935.8	0.00	99.99
49.0-50.0	1.8	0.1	4935.9	0.00	100.00
50.0-51.0	1.1	0.1	4936.0	0.00	100.00
51.0-52.0	0.7	0.1	4936.1	0.00	100.00
52.0-53.0	0.0	0.0	4936.1	0.00	100.00
53.0-54.0	0.0	0.0	4936.1	0.00	100.00
54.0-55.0	0.0	0.0	4936.1	0.00	100.00
55.0-56.0	0.0	0.0	4936.1	0.00	100.00
56.0-57.0	0.0	0.0	4936.1	0.00	100.00
57.0-58.0	0.0	0.0	4936.1	0.00	100.00
58.0-59.0	0.0	0.0	4936.1	0.00	100.00
59.0-60.0	0.0	0.0	4936.1	0.00	100.00
60.0-61.0	0.0	0.0	4936.1	0.00	100.00
61.0-62.0	0.0	0.0	4936.1	0.00	100.00
62.0-63.0	0.0	0.0	4936.1	0.00	100.00
63.0-64.0	0.0	0.0	4936.1	0.00	100.00
64.0-65.0	0.0	0.0	4936.1	0.00	100.00
65.0-66.0	0.0	0.0	4936.1	0.00	100.00
66.0-67.0	0.0	0.0	4936.1	0.00	100.00
67.0-68.0	0.0	0.0	4936.1	0.00	100.00
68.0-69.0	0.0	0.0	4936.1	0.00	100.00
69.0-70.0	0.0	0.0	4936.1	0.00	100.00
70.0-71.0	0.0	0.0	4936.1	0.00	100.00
71.0-72.0	0.0	0.0	4936.1	0.00	100.00

C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Candlepower Table

Unit: cd

G\C	C0.0	C180.0								
G0.0	19547.9	19547.9								
G1.0	19504.4	19512.8								
G2.0	19347.1	19379.4								
G3.0	19012.9	19057.8								
G4.0	18393.6	18430.1								
G5.0	17507.4	17507.4								
G6.0	16462.7	16434.6								
G7.0	15312.5	15221.2								
G8.0	14052.9	13916.7								
G9.0	12720.2	12525.0								
G10.0	11355.2	11117.9								
G11.0	10033.8	9769.7								
G12.0	8782.5	8521.3								
G13.0	7667.5	7367.0								
G14.0	6688.7	6374.1								
G15.0	5848.9	5554.0								
G16.0	5142.6	4874.3								
G17.0	4541.5	4309.8								
G18.0	4024.7	3854.8								
G19.0	3593.6	3460.2								
G20.0	3248.1	3149.9								
G21.0	2968.7	2897.1								
G22.0	2738.4	2686.4								
G23.0	2544.6	2508.1								
G24.0	2388.7	2359.2								
G25.0	2255.3	2231.4								
G26.0	2138.8	2121.9								
G27.0	2043.3	2029.2								
G28.0	1960.4	1946.4								
G29.0	1884.6	1873.3								
G30.0	1815.8	1787.7								
G31.0	1723.1	1676.7								
G32.0	1606.5	1546.1								
G33.0	1474.5	1402.9								
G34.0	1336.9	1254.0								
G35.0	1183.8	1103.8								
G36.0	1036.4	956.3								

C Plane (°):0.0-180.0: 180.0
Test Lab:
Test Type: TYPE C
Temperature:
Operator:

Gamma Plane (°):0.0-90.0:1.0
Test Device: CHL-6E
Distance: 2.905 m
Humidity:
Inspector:

Candlepower Table (Continue 1)

Unit: cd

G\C	C0.0	C180.0								
G37.0	893.1	808.9								
G38.0	748.5	662.8								
G39.0	605.3	518.2								
G40.0	466.2	389.0								
G41.0	337.0	269.6								
G42.0	227.5	168.5								
G43.0	134.8	87.1								
G44.0	68.8	44.9								
G45.0	39.3	28.1								
G46.0	25.3	11.2								
G47.0	9.8	2.8								
G48.0	4.2	1.4								
G49.0	4.2	1.4								
G50.0	1.4	0.0								
G51.0	2.8	0.0								
G52.0	0.0	0.0								
G53.0	0.0	0.0								
G54.0	0.0	0.0								
G55.0	0.0	0.0								
G56.0	0.0	0.0								
G57.0	0.0	0.0								
G58.0	0.0	0.0								
G59.0	0.0	0.0								
G60.0	0.0	0.0								
G61.0	0.0	0.0								
G62.0	0.0	0.0								
G63.0	0.0	0.0								
G64.0	0.0	0.0								
G65.0	0.0	0.0								
G66.0	0.0	0.0								
G67.0	0.0	0.0								
G68.0	0.0	0.0								
G69.0	0.0	0.0								
G70.0	0.0	0.0								
G71.0	0.0	0.0								
G72.0	0.0	0.0								
G73.0	0.0	0.0								

C Plane (°):0.0-180.0: 180.0

Test Lab:

Test Type: TYPE C

Temperature:

Operator:

Gamma Plane (°):0.0-90.0:1.0

Test Device: CHL-6E

Distance: 2.905 m

Humidity:

Inspector:

Unit: cd

Inspector: